

Republic of the Philippines
Jose R. Reyes Memorial Medical Center
(02) 711-9491 loc. 245 / 8562-5338

MINUTES OF THE MEETING
PRE-BIDDING CONFERENCE OF VARIOUS MEDICAL EQUIPMENT
16 June 2022

Present Were:

HBAC-A2 Chairperson:

1. Dr. Emelito O. Valdez – Tan

HBAC-A2 Vice-Chairperson:

2. Ms. Natividad Sarmiento

HBAC-A2 Members:

3. Dr. Juanita Castillo - Cariño
4. Ms. Joseph T. Gatchalian
5. Mr. Arlie Felvin Gacias

Technical Working Group

1. Ligasure Hand Instrument – Dr. Rufino T. Agudera, Dr. Juan Emmanuel T. Montaña, Dr. Jessie L. Harina

AHW Representative

1. Mr. Arnold Pascual

Bidder/s:

1. Sahar L. Gregorio – Zafire Distributor
2. Joan Palma – RBGM Medical Express
3. Clair Ernesto – Health Rush Enterprises
4. Jennifer Telen – Dynamed Healthcare
5. Monica Santillan – Johnson & Johnson
6. Enrico Arellano – Icarelife
7. Roel Sibayan – Ritegroup
8. Xavier Natividad – Grand Lime
9. Kietshand May V. Dela Cruz – R & A Dental Supply Center Co.

The meeting of the Bids and Awards Committee (BAC), held at Derma Conference Room was called to order at 3:17 pm, of 16 June 2022 and was presided by the BAC Chairman, Dr. Emelito O. Valdez – Tan.

I. Business Matter

- a) The Schedule of Requirements, Approved Budget for the Contract (ABC) and Delivery days was flashed in the screen and reviewed by the Bidder/s.
- b) Bidder/s were reminded to itemized costing of each item offered (including AVR, Accessories & Installation (if included in the specification). A separate sheet to be included in the Financial Proposal was recommended, no format was required as long as indicates the costing of each item offered.

LIGASURE HAND INSTRUMENT

- The following specifications were discussed and clarified by the Bidder/s with BAC & TWG – End-user, suggested the following changes or be remained for clarification to which the BAC agreed and will be reflected with the bid bulletin, if granted / finalize:
 - For project title “LIGASURE” was requested to be deleted and was changed to “ENERGY BASED VESSEL SEALING DEVICE”.
 - “Power Supply” and “60Hz” was added.
 - In Max VA nominal line Voltage, “Volt-Ampere” was added.
 - In Max Mains Current, “Amperes” was added.
 - In Max Heat Dissipation, “Watts” was added.
 - In Fuses, “Amperes” was added.
 - In Power Cord, “or its equivalent” was added.
 - Under MARYLAND JAW LAP DEVICE, Shaft length was change to “at least 35cm”
 - Under BLUNT TIP LAP SEALER DIVIDER, Shaft length was change to “at least 35cm”
 - Under MARYLAND JAW OPEN DEVICES, Jaw angle was change to “at least 20 Deg”
 - For Bidder’s certificate that the BRAND must be in the Philippines, was change to “2012”.

PATIENT MONITOR WITH CENTRAL STATION MONITORING SYSTEM INCLUDING SUPPLIES AND INSTALLATION


- The following specifications were discussed and clarified by the Bidder/s, suggested the following changes or be remained for clarification to which the BAC agreed and will be reflected with the bid bulletin, if granted / finalize:
 - At least “Lead selection switch” was clarified if can be “*Or its equivalent*”.
 - At least “2.5 to 20 mm/mV” was clarified if can be in “*cm*”.
 - For “Low and high pass signal” was clarified if can be in “*3 modes / terms: diagnostic, monitoring and operating*”.
 - In ECG recorder/printer capable of printing “at least 4 waveforms” if can be “*at least 3 waveforms*”.
 - “Digital thermometer temperature probe” was clarified “*if same as skin probe*”.
 - In Systolic and diastolic blood pressure measurement with a maximum pressure reading of “300mmHg” was clarified if can be “*275mmHg*”.
 - In cabinet cart with “4 anti-static and rust-free swivel wheels” was clarified “*if can be 5 wheels or be at least 4 wheels*”.
 - Protective case was clarified if this is a nylon dust-cover.


There, having no other remaining topics for discussion, the meeting adjourned at 4:41pm.

Prepared by:


MADRIELE F. DASMARIÑAS
BAC Secretariat

Noted by:


EMELITO VALDEZ – TAN, MD
Chairperson – HBAC – A2



at
i/23
CHEL/mau