



JOSE R. REYES MEMORIAL MEDICAL CENTER
Rizal Avenue, Sta. Cruz, Manila
BAC SECRETARIAT

SEALED BID

SEALED BID NO.: HBAC-A2-2021-04-543

DATE: April 2021

Gentlemen:

Please quote your price for item/s indicated below. Price quoted is final, and without condition and inclusive of all taxes. Bid Proposal should be submitted in sealed envelope, properly labelled, indicating the date and Sealed Bid No., not later than **8:30 am on April 2021** at the BAC Secretariat Office and thereafter publicly opened. Any change/erasure must be duly initialed otherwise, the bid shall be disqualified. Attendance of authorized representative is not mandatory on bid opening, in accordance with **Section 53.2 of the Revised Implementing Rules and Regulation of the Republic Act 9184.**

THE COMMITTEE RESERVES THE RIGHT TO REJECT ANY AND ALL BIDS, DECLARE A FAILURE OF BIDDING AT ANY TIME PRIOR TO THE CONTRACT AWARD, OR NOT TO AWARD THE CONTRACT, WITHOUT THEREBY INCURRING ANY LIABILITY, AND MAKE NO ASSURANCE THAT A CONTRACT SHALL BE ENTERED INTO AS A RESULT OF THE BIDDING.

For:
EMELITO VALDEZ-TAN, M.D.
Chairman, HBAC-A-2

PR No.: 2021-03-1469 Medical Service

Qty.	Unit	Description/Particulars	ABC	Unit Price	Total Price
10	Unit	<p>Nasal High Flow Oxygen Therapy Device Branded/Brand New</p> <p>Specification:</p> <p>Clinical Features</p> <ul style="list-style-type: none"> ➤ Temperature Settings: 31-37°C (minimum range) ➤ Max Temperature: 43°C ➤ Humidity Requirement: >12->33 mg/L (10-95% Relative Humidity) ➤ Flow Range: 2 - 60 Lpm (minimum range) ➤ Maximum Oxygen Input: 60 Lpm ➤ FiO2 Monitoring Range: 21 - 100% ➤ FiO2 Monitoring Accuracy: +/- 4% ➤ Warm - up Time: at least 10 - 30mins (at temp range 31 - 37°C) ➤ Auditory Alarms: >/= 45 dbA; ➤ Visual Alarms: can detect: unplugged breathing tube; leak in the system; blockage in the system; if measured oxygen level has fallen below allowed limit; if measured oxygen has exceeded the allowed limit; if unit cannot reach the target flow setting; if unit cannot reach target temperature 	375,000.00		



JOSE R. REYES MEMORIAL MEDICAL CENTER
 Rizal Avenue, Sta. Cruz, Manila
BAC SECRETARIAT

SEALED BID

- setting; if unit has been disconnected from the main power supply
- Alarm Silence: at least 1 min
- Disinfection kit: with Disinfection Filter, Sponge
- Oxygen analyzer accuracy: \pm (2.5% + 2.5% of gas level)

Physical

- Weight: less than 10 lbs
- Mobility: with hospital stand and at least 4 wheels, and pole mounting tray

Electrical

- Frequency: 50 – 60Hz
- Current/Voltage Requirement: Autovolt (110 – 240)
- UPS (Uninterruptible Power Supply: at least 30mins to 1 hour)

Accessories

- Heated Non-condensing Breathing Tubes/Circuit: Disposable
- Humidification Chamber: Disposable
- HF Nasal Cannula Patient Interface: Small, Medium, Large, Tracheostomy connection adapter, mask interface adapter
- Air Filter: 3 months maximum period of use

Miscellaneous

- Oxygen inlet extension kit Hospital stand Pole mounting tray: with wheels
- Mobility: with hospital stand at least 4 wheels, and pole mounting tray

I. Standard Requirements:

- Current and Valid Certificate of Manufacturer's compliance with ISO Certified or its equivalent.
- Bidder's certificate that the BRAND must be in the Philippines since 2011
- User's Manual in English Language (2 hard copies and 1 soft copy) upon delivery
- Service/Technical and Maintenance Manual (2 hard copies and 1 soft copy) upon delivery



JOSE R. REYES MEMORIAL MEDICAL CENTER

Rizal Avenue, Sta. Cruz, Manila

BAC SECRETARIAT

SEALED BID

		<ul style="list-style-type: none"> ▪ Latest Certificate of Traceability from the Manufacturer ▪ Proposed Costing of Preventive Maintenance and Calibration Program for sophisticated equipment, consumables/ accessories. ▪ Printing and/or Etching of DOH official logo/ letters (if possible "JRRMMC" in all equipment) in conspicuous space of the equipment but will not affect its function. <p>II. Training Requirements</p> <ul style="list-style-type: none"> ▪ Training at least two (2) from end-users and one (1) from engineering <p>III. After Sales Service</p> <ul style="list-style-type: none"> ▪ Bidder's certificate that parts shall be available at the authorized Philippine service center /s for a period of five (5) years after the warranty period. <p>IV. Warranty Certificate</p> <ul style="list-style-type: none"> ▪ Minimum of one (1) year on parts and service. Warranty Certificate for parts and service, upon delivery, inspection and acceptance ▪ The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. <p>V. Preventive Maintenance and Calibration</p> <p>Free Quarterly preventive Maintenance and Calibration within the warranty period</p>			
--	--	--	--	--	--

***Kindly provide the following documentary requirements (Mayor's/Business Permit, PhilGEPS certificate, Income/Business Tax Return and Omnibus Sworn Statement)**

Please check Mode of Payment:

___1. Terms

Name of Company

___2. Cheque upon Delivery (COD)
No Clearing time before delivery

Address/TIN No.

Authorized signature over printed name/ Tel No.:

NOTE: Signing for these request for quotation, is conforming to our payment terms you check